

Catheter tipping, also known as end forming, is the process of shaping the end of a thermoplastic tube using a thermally conductive mold. For medical device manufacturers, there are several common methods for application of energy to the conductive mold including: Hot air processing. Induction heating

From:

<https://neurosurgerywiki.com/wiki/> - **Neurosurgery Wiki**

Permanent link:

https://neurosurgerywiki.com/wiki/doku.php?id=catheter_tip

Last update: **2024/06/07 02:58**

