

(SPGR) The SPGR pulse sequence is similar to the spoiled [GRASS](#) sequence. The spoiled gradient recalled (SPGR) acquisition in steady state uses semi-random changes in the phase of the radio frequency (RF) pulses to produce a spatially independent phase shift.

From:

<https://neurosurgerywiki.com/wiki/> - **Neurosurgery Wiki**

Permanent link:

[https://neurosurgerywiki.com/wiki/doku.php?id=gradient\\_recalled](https://neurosurgerywiki.com/wiki/doku.php?id=gradient_recalled)

Last update: **2024/06/07 02:53**

