

van Doormaal et al., evaluated 9 commonly used dural sealants, including **Tachosil** (Takeda Inc, Osaka, Japan), **Adherus** (Hyperbranch Inc, Durham, North Carolina), **Duraform** (Codman, Raynham, Massachusetts), **Tissudura** (Baxter, Deerfield, Illinois), **Hemopatch** (Baxter), **TissuePatchDural** (**Tissuemed**, Leeds, United Kingdom), **Tisseel** (Baxter), **Duragen** Secure (**Integra**, Plainsboro, New Jersey), and **Duraseal**, (Integra). Sealants were tested in 2 novel in Vitro setups using fresh porcine dura: the first tested the acute burst pressure of a sealed 3-mm gap, while the second examined resistance to a pressure wave mimicking intracranial pressure for 72 h.

**Adherus** showed the highest mean burst pressure ( $87 \pm 47$  mmHg) followed by Tachosil ( $71 \pm 16$  mmHg) and Duraseal ( $51 \pm 42$  mmHg); these were the only 3 sealants showing burst pressures above normal physiological intracranial pressure. In the 72-h setup, only Adherus and Duraseal maintained appropriate sealing for the duration of the experiment. Tachosil released from the dura after 1.4 h (95% confidence interval, -1.8-4.7).

Given the high cost of sealants and the results of this study, they advocate a critical attitude toward sealant application as an adjunct to classic **dural closure**<sup>1)</sup>.

<sup>1)</sup>

van Doormaal T, Kinaci A, van Thoor S, Redegeld S, Bergmann W, van der Zwan A. Usefulness of Sealants for Dural Closure: Evaluation in an In Vitro Model. Oper Neurosurg (Hagerstown). 2018 Oct 1;15(4):425-432. doi: 10.1093/ons/oxz260. PubMed PMID: 29281065.

From:  
<https://neurosurgerywiki.com/wiki/> - **Neurosurgery Wiki**



Permanent link:  
<https://neurosurgerywiki.com/wiki/doku.php?id=hyperbranch>

Last update: **2024/06/07 02:55**