

Initially, each electrode contact on the lead is tested in a monopolar configuration with the electrode as negative (cathode) and the neurostimulator case as positive (anode), a process referred to as monopolar review.

From:

<https://neurosurgerywiki.com/wiki/> - **Neurosurgery Wiki**

Permanent link:

[https://neurosurgerywiki.com/wiki/doku.php?id=monopolar\\_review](https://neurosurgerywiki.com/wiki/doku.php?id=monopolar_review)

Last update: **2024/06/07 02:54**

