

Carotid stents of different design and configuration are available. Depending on the density of struts, stents can be classified into stents with a closed-cell or an open-cell configuration.

Closed-cell stents are characterized by small free cell areas between struts, whereas open-cell stents have larger gaps uncovered

From:

<https://neurosurgerywiki.com/wiki/> - **Neurosurgery Wiki**

Permanent link:

https://neurosurgerywiki.com/wiki/doku.php?id=open-cell_stent

Last update: **2024/06/07 02:54**

