

# p48 MW

[https://www.phenox.net/fileadmin/uploads/Brochures/p48\\_KIF-0057A\\_HR.pdf](https://www.phenox.net/fileadmin/uploads/Brochures/p48_KIF-0057A_HR.pdf)

**Flow diverter** (FD) is increasingly used in the management of **wide-neck intracranial aneurysms**. Despite a reported lower efficacy in **middle cerebral artery aneurysms**, they are still being utilized. **Microsurgery** is best considered as an index treatment, but can also be a safe and effective treatment when encountering a persistent MCA aneurysm after prior FD. As there is a paucity of literature and more cases of failed FD are expected to appear, Veken et al. want to add the experience to the existing literature. The microsurgical management of a persistent MCA bifurcation aneurysm, 3 years after a **p48 MW** HPC Flow Diverter (**phenox** GmbH, Bochum Germany) insertion is reported, and the relevant literature discussed <sup>1)</sup>

<sup>1)</sup>

Veken JV, Keukeleire K. Clipping of a persistent middle cerebral artery aneurysm after previous flow diverter placement: An illustrative case and review of the literature. J Cerebrovasc Endovasc Neurosurg. 2024 Jan 31. doi: 10.7461/jcen.2024.E2023.09.002. Epub ahead of print. PMID: 38291255.

From:

<https://neurosurgerywiki.com/wiki/> - **Neurosurgery Wiki**

Permanent link:

[https://neurosurgerywiki.com/wiki/doku.php?id=p48\\_mw](https://neurosurgerywiki.com/wiki/doku.php?id=p48_mw)

Last update: **2024/06/07 02:57**

