

The noninvasive device based on a modified Penaz technique cannot replace arterial blood pressure monitoring or TEE in anesthetized patients undergoing neurosurgery in sitting position ¹⁾.

¹⁾

Schramm P, Tzanova I, Gööck T, Hagen F, Schmidtman I, Engelhard K, Pestel G. Noninvasive Hemodynamic Measurements During Neurosurgical Procedures in Sitting Position. J Neurosurg Anesthesiol. 2016 Mar 18. [Epub ahead of print] PubMed PMID: 26998648.

From:

<https://neurosurgerywiki.com/wiki/> - **Neurosurgery Wiki**

Permanent link:

https://neurosurgerywiki.com/wiki/doku.php?id=penaz_technique

Last update: **2024/06/07 02:58**

