

RF pulses deposit RF energy that can cause unwanted heating of the patient. This heating is measured by the specific absorption rate (SAR) (in watts per kilogram).

From:

<https://neurosurgerywiki.com/wiki/> - **Neurosurgery Wiki**

Permanent link:

https://neurosurgerywiki.com/wiki/doku.php?id=radiofrequency_pulse

Last update: **2024/06/07 02:56**

