

Siphon effect

The [siphon](#) effect may be created by the elevation of the [ventricular catheter](#) with respect to the distal [catheter](#) (i.e., when the patient sits, stands, or is held erect).

From:

<https://neurosurgerywiki.com/wiki/> - **Neurosurgery Wiki**

Permanent link:

https://neurosurgerywiki.com/wiki/doku.php?id=siphon_effect

Last update: **2024/06/07 03:00**

