

In [spinal oncology](#), [titanium](#) implants pose several challenges including [artifacts](#) on advanced imaging and therapeutic radiation perturbation. To mitigate these effects, there has been increased interest in radiolucent [carbon fiber](#) (CF) and CF-reinforced [polyetheretherketone](#) (CFR-PEEK) [implants](#) as an alternative for [spinal reconstruction](#).

From:

<https://neurosurgerywiki.com/wiki/> - **Neurosurgery Wiki**

Permanent link:

https://neurosurgerywiki.com/wiki/doku.php?id=spinal_reconstruction

Last update: **2024/06/07 02:50**

