

Ventriculoscope

LOTTA® System Two ventriculoscopes – diameter 3.6 mm and 6.1 mm – for the entire range of endoscopic intracranial procedures

6° angle of view for visual control of instruments after exiting the working channel

Specially developed instrument sets for LOTTA® and Little LOTTA®

From:

<https://neurosurgerywiki.com/wiki/> - **Neurosurgery Wiki**

Permanent link:

<https://neurosurgerywiki.com/wiki/doku.php?id=ventriculoscope>

Last update: **2024/06/07 02:58**

